

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 4489 (1968): Forceps, Nasal Turbinate, Luc's [MHD 4: Ear, Nose and Throat Surgery Instruments]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



IS : 4489 - 1968

Indian Standard
SPECIFICATION FOR
FORCEPS, NASAL TURBINATE, LUC'S
(First Reprint JUNI 1990)

UDC 615.472.41 : 616.211 - 089

© Copyright 1968

BUREAU OF INDIAN STANDARDS
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

Gr 2

1968

Indian Standard

SPECIFICATION FOR FORCEPS, NASAL TURBINATE, LUC'S

Surgical Instruments Sectional Committee, CPDC 11

Chairman

COL R. D. AYYAR

Representing

Safdarjang Hospital, Central Government Health Scheme (Ministry of Health & Family Planning), New Delhi

Members

CHIEF SALES MANAGER

WORKS MANAGER (*Alternate*)

SHRI R. D. GUPTA

DR (KUMARI) S. D. GUPTA

BRIG IQBAL CHAND

SHRI G. B. JAKHETIA

SHRI V. KRIHNAMOORTHY

SHRI SANTI KUMAR MUKERJI

SHRI M. G. PANDIT

DR A. RAMDAS

SHRI S. SEN GUPTA

SHRI MAHESH M. SHAH

MAJ SOHAN SINGH

SHRI K. L. TALWAR

SHRI A. N. VIO

SHRI R. N. VIO (*Alternate*)

SHRI A. B. RAO,

Director / Condr Prod) (*Secretary*)

Indian Drugs & Pharmaceuticals Ltd, New Delhi

National Metallurgical Laboratory (CSIR), Jamshedpur

Lady Hardinge Medical College & Hospital, New Delhi

Directorate General of Armed Forces Medical Services (Ministry of Defence)

Development Commissioner, Small Scale Industries

Directorate General of Technical Development

H. Mukerji & Banerjee Surgical Pvt Ltd, Calcutta

Ministry of Health & Family Planning

Ministry of Railways

Baruipur Surgical Instrument Centre (Directorate of Industries, Government of West Bengal)

Shah Medical & Surgical Co Ltd, Baroda

Directorate General of Inspection (Ministry of Defence)

Chief Inspectorate of General Stores [Ministry of Defence (DGI)], Kanpur

ISCOMED Industries Private Ltd, Bombay

Director General, ISI (*Ex-officio Member*)

(Continued on page 2)

BUREAU OF INDIAN STANDARDS
MANAK BHAYAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

IS : 4489 - 1968

(Continued from page 1)

Precision Type Surgical Instruments Subcommittee, CPDC 11 : 5

Convener

DR J. L. GUPTA

Representing

Safdarjang Hospital, Central Government Health Scheme (Ministry of Health & Family Planning), New Delhi

Members

DR JACOB CHANDI

Christian Medical College, Vellore

DR G. S. GOPALAN

Jaya Surgicals, Madras

SHRI G. JAYA PRASAD (Alternate)

DR GOPINATH

All India Institute of Medical Sciences, New Delhi

BRIG IQBAL CHAND

Directorate General of Armed Forces Medical Services (Ministry of Defence)

SHRI SANTI KUMAR MUKERJI

H. Mukerji & Banerjee Surgical Pvt Ltd, Calcutta

SHRI H. C. MUKHERJEA

Chief Inspectorate of General Stores [Ministry of Defence (DGI)], Kanpur

SHRI J. PRABHAKAR

Indian Drugs & Pharmaceuticals Ltd, Madras

DR T. T. RAMALINGAM

In personal capacity (4-A Kilpauk Garden Road, Kilpauk, Madras)

DR P. B. RAO

Jawaharlal Institute of Post-Graduate Medical Education & Research (Ministry of Health & Family Planning), Pondicherry

Indian Standard

SPECIFICATION FOR FORCEPS, NASAL TURBINATE, LUC'S

0. FOREWORD

0.1 This Indian Standard was adopted by the Indian Standards Institution on 29 January 1968, after the draft finalized by the Surgical Instruments Sectional Committee had been approved by the Consumer Products Division Council.

0.2 The formulation of Indian Standards on surgical instruments has been taken up on the recommendations of the Advisory Committee for Development of Surgical Instruments, Equipment and Appliances, Government of India.

0.3 This standard is one of a series of Indian Standards on surgical instruments.

0.4 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS : 2-1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard lays down the requirements for Luc's nasal turbinate forceps.

2. MATERIAL

2.1 The forceps shall be made of stainless steel conforming to Designation 30Cr13 of Schedule V of IS : 1570-1961†.


2.2 The screw shall conform to the Designation mentioned in 2.1 or 20Cr18Ni2 of Schedule V of IS : 1570-1961†.

3. SHAPE AND DIMENSIONS

3.1 The general shape and dimensions of the forceps shall be in accordance with Fig. 1 and 2.


*Rules for rounding off numerical values.

†Schedules for wrought steels for general engineering purposes.



*See IS : 3642-1966 'General requirements for surgical instruments'.
All dimensions in millimetres.

FIG. 1 FORCEPS, NASAL TURBINATE, LUC'S PATTERN



All dimensions in millimetres.

**FIG. 2 SHAPES OF JAWS OF FORCEPS, NASAL TURBINATE,
LUC'S PATTERN**

3.2 Joints — The joint shall be of the cut-through type. The other requirements for the joint shall conform to 6 of Section 2 of IS : 3642-1966*.

4. WORKMANSHIP AND FINISH

4.1 The forceps shall be of a symmetrical and balanced construction and shall open and close with even jerkless movement. The jaws shall be semi-sharp over the periphery as shown in Fig. 2 and shall engage accurately in the closed position. The arms of the forceps shall be free from surface defects, such as scales, burrs, pits, cracks and seams. All sharp edges, except the periphery of the jaws, shall be rounded smooth. The forceps shall be passivated and polished bright.

5. HEAT-TREATMENT

5.1 The forceps shall be suitably hardened and tempered to give a uniform hardness of 430 to 490 HV.

6. TESTS

6.1 Performance — The jaws of the forceps shall be made to bite on a double layer of wet blotting paper. The jaws shall leave a clear impression of the complete semi-sharp edge on the blotting paper. They shall also be similarly tested on a flat 3 mm thick cork or cork sheeting and shall leave a clear impression of the complete semi-sharp edge on the cork.

*General requirements for surgical instruments.

6.2 Flexibility — The jaws of the forceps shall be closed on a 3 mm thick hard rubber piece and the arms pressed fully. No permanent set shall be noticed on the forceps and after this test the forceps shall satisfy the performance test prescribed in 6.1.

6.3 Corrosion Resistance — The forceps shall be tested for corrosion resistance as given in 6.3.1.

6.3.1 Copper Sulphate Test — The sample shall be scrubbed with soap and warm water, rinsed in hot water, followed by a dip in ethyl alcohol (95 percent) and dried. The sample shall be completely immersed in copper sulphate solution at room temperature for six minutes and shall then be washed off with fresh water or wet cotton wool. The copper sulphate solution shall be made up as follows:

Copper sulphate ($\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$)	4.0 g
Sulphuric acid (H_2SO_4) (sp gr 1.84)	10.0 g
Water (H_2O)	90.0 ml

There shall be no red stains or spots on the sample after the test, but the polished surface getting dull may be permitted.

7. MARKING

7.1 Each forceps shall be clearly and indelibly marked with the manufacturer's name, initials or trade-mark, in addition to the words 'STAINLESS STEEL'.

7.1.1 The forceps may also be marked with the ISI Certification Mark.

NOTE — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act, and the Rules and Regulations made thereunder. Presence of this mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard, under a well-defined system of inspection, testing and quality control during production. This system, which is devised and supervised by ISI and operated by the producer, has the further safeguard that the products as actually marketed are continuously checked by ISI for conformity to the standard. Details of conditions, under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

8. PACKING

8.1 The forceps shall be wrapped in moisture-proof paper or polyethylene bag; then they shall be packed in a manner that coming into contact with each other is avoided.

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002

Telephones: 331 01 31, 331 13 75

Telegrams: Manaksanstha
(Common to all Offices)

Regional Offices:

Telephone

Central	Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002	{ 331 01 31 331 13 75
*Eastern	1/14 C. I. T. Scheme VII M, V. I. P. Road, Manikotla, CALCUTTA 700054	36 24 99
Northern	SCO 445-446, Sector 35-C, CHANDIGARH 160036	{ 2 18 43 3 16 41 41 24 42
Southern	C. I. T. Campus, MADRAS 600113	{ 41 25 19 41 29 16
†Western	Manakalaya, E9 MIDC, Marol, Andheri (East), BOMBAY 400093	6 32 92 95

Branch Offices:

'Pushpak', Nurmohamed Shaikh Marg, Khanpur, AHMADABAD 380001	{ 2 63 48 2 63 49
†Peenya Industrial Area 1st Stage, Bangalore Tumkur Road BANGALORE 560058	{ 38 49 55 38 49 56
Gangotri Complex, 5th Floor, Bhambhada Road, T. T. Nagar, BHOPAL 462003	6 67 16
Plot No. 82/83, Lewis Road, BHUBANESHWAR 751002	5 36 27
53/5, Ward No. 29, R.G. Barua Road, 5th Byelane, GUWAHATI 781003	3 31 77
5-8-56C L. N. Gupta Marg (Nampally Station Road), HYDERABAD 500001	23 10 83
R14 Yudhister Marg, C Scheme, JAIPUR 302005	{ 6 34 71 6 98 32
117/418 B Sarvodaya Nagar, KANPUR 208005	{ 21 68 76 21 82 92
Patliputra Industrial Estate, PATNA 800013	6 23 05
T.C. No. 14/1421, University P.O., Palayam TRIVANDRUM 695035	{ 6 21 04 6 21 17

Inspection Offices (With Sale Point):

Pushpanjali, First Floor, 205-A West High Court Road, Shankar Nagar Square, NAGPUR 440010	2 51 71
Institution of Engineers (India) Building, 1332 Shivaji Nagar, PUNE 411005	5 24 35

*Sales Office in Calcutta is at 5 Chowringhee Approach, P. O. Princep Street, Calcutta 700072

†Sales Office in Bombay is at Novati Chambers, Grant Road, 89 65 28
Bombay 400007

†Sales Office in Bangalore is at Unity Building, Narasimharaja Square, 22 36 71
Bangalore 560002

Reprography Unit, BIS, New Delhi, India

AMENDMENT NO. 1 JANUARY 1979
TO
IS : 4489-1968 SPECIFICATION FOR
FORCEPS, NASAL TURBINATE, LUC'S

Alteration

(*Page 6, clauses 6.3 and 6.3.1*) — Substitute the following for the existing clauses:

'6.3 Corrosion Resistance — The instrument shall satisfy the boiling and autoclaving test as specified in IS : 7531-1975*.'

Addendum

(*Page 6, foot-note*) — Add the following new foot-note at the bottom of the page:

* *Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments.'

(CPDG 26)

Reprography Unit, BIS, New Delhi, India;